Jackson Spinal Surgical Tabletop
Models: 5928, 5872, 5972, 5824, 5924, 5843, 5943, 5943ACP
of the Modular Table System (MTS) product line

Work Instruction Number 50603-07

Contact Information:

Date: _____Model/Serial#: _____ Account #: _____

Hospital Name: _____ City & State: _____

Contact Name: _____ Position: _____ Phone: _____

Equipment Location: _____

Service Representative: _____ Service Organization: _____

Service Rep. phone: _____ Home location: _____

Customer contact accepting service: _____ _____ _____

Name please print Signature Date

Service Representative: _____ _____ _____

Name please print Signature Date

Using this report: Please proceed through each step and place an "x" in the "( )" as each item representing **Pass Fail** **( ) ( )** criteria is inspected. If a step has failed, please indicate with an "x" and note the reasons in the comments section. If you resolve the failure place an "x" in the "Pass" area and indicate the corrective action again in the comments section. If you are unable to perform any Pass/Fail check, please contact Mizuhosi Technical Services for assistance and circle the check

Biohazards when servicing equipment: Please take appropriate precautions for servicing potentially contaminated biohazard equipment or working in biohazard surroundings. Please wear protective surgical gloves and log sleeved gowns at all times. If your gloves or protective clothing tears, replace immediately. Have the hospital staff clean the equipment appropriately prior to servicing. Clean your tools that may contain biohazards prior to placing back in your tool kit for removal from the location. If you suspect biohazards are present do not service the equipment until they have been cleaned.

(If serviced by an outside service organization, the record must be reviewed, signed, and dated by Technical Services Dept. Refer to last page of this document).

Surgical Tops covered:

5828 Jackson Spinal Surgery Top, 350lb

5928 Jackson Spinal Surgery Top, 500lb

5872 Jackson Spinal Surgery Top with ShereGuard Hip and Thigh Gel Pads, 350lb

5972 Jackson Spinal Surgery Top with ShereGuard Hip and Thigh Gel Pads, 500lb

5824 Jackson Spinal Surgery Top with Gimbals and Pads, 350lb

5924 Jackson Spinal Surgery Top with Gimbals and Pads, 500lb

5843 Jackson Spinal Surgery Top with Gimbals and Pads, 350lb

5943 Jackson Spinal Surgery Top with Gimbals and Pads, 500lb

Tools Required:

Lubricant, silicon, spray

Grease, Lithium bases, white, multi-purpose

Materials Required:

Jackson Spinal Surgery & Imaging Table Tops Owner's Manual – NW0504

Pass Fail

1. Top Inspection and fit.

1.1 ☐ ☐ Table top is clean, without signs of hazardous contamination, is free of wear, cracks, delamination or splinters and its consistent with proper care.

Comments: _____

1.2 ☐ ☐ All labels are intact and legible.

Comments: _____

1.3 ☐ ☐ Gimbals are secure, rotate to stops and return to center position.
Note: earlier models may not be equipped with gimbals.

Comments: _____

1.4 ☐ ☐ Fits in all positions on removable H-frames, T-pins slide into locked position freely.

Comments: _____

2. Standard Attachments inspect if its available.

2.1 ☐ ☐ Head Support fits firmly onto Jackson frame and its free of visual cracks or breaks.

Comments: _____

2.2 ☐ ☐ Armboard(s) fits firmly onto Jackson frame and its free of visual cracks, breaks or puncture.
Commments: _____

2.3 ☐ ☐ Chest pad fits firmly onto Jackson frame and its free of visual cracks, breaks or puncture.
Comments: _____

2.4 ☐ ☐ Hip and Thigh Pads fits firmly onto Jackson frame and its free of visual cracks, breaks or puncture.
Comments: _____

2.5 ☐ ☐ Leg Support(s) fits firmly onto Jackson frame and its free of visual cracks, breaks or puncture.
Comments: _____

2.6 ☐ ☐ Thoracic, Buttock and Leg Straps are free of damage and attach securely.
Comments: _____

This section is for Technical Services use only.

Reviewed**:** _____ Signature: _____ Date: _____